

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091224\
 Data File : VU060725.D
 Acq On : 12 Sep 2024 19:37
 Operator : MD/SY
 Sample : P3918-17
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AP8

Quant Time: Sep 13 05:50:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Aug 17 05:03:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	8442	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	12361	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	3669	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	6017	0.438	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery =	88.000%		
4) 1,2-Dichloroethane-d4	5.685	65	3649	0.501	ug/L	-0.02
Spiked Amount 0.500	Range 70 - 130		Recovery =	100.000%		
7) 1,2-Dichloropropane-d6	6.687	67	3893	0.507	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery =	102.000%		
8) Toluene-d8	7.888	98	8403	0.529	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery =	106.000%		
10) 1,1,2,2-Tetrachloroeth...	10.745	84	3211	0.477	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery =	96.000%		
Target Compounds					Qvalue	
6) Trichloroethene	6.544	95	1102	0.141	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091224\
Data File : VU060725.D
Acq On : 12 Sep 2024 19:37
Operator : MD/SY
Sample : P3918-17
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AP8

Quant Time: Sep 13 05:50:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM081624.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Aug 17 05:03:57 2024
Response via : Initial Calibration

