

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091323\
 Data File : VU055286.D
 Acq On : 13 Sep 2023 15:52
 Operator : MD/SY
 Sample : 04350-07DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH4M8DL

Quant Time: Sep 14 05:05:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 14 05:00:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	175824	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	196455	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	93372	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	26016	3.157	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	63.200%	
7) Chloroethane-d5	1.907	69	44711	4.687	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	93.800%	
11) 1,1-Dichloroethene-d2	2.560	65	14600	3.109	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	62.200%	
20) 2-Butanone-d5	4.631	46	170579	71.332	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	142.660%#	
24) Chloroform-d	5.062	84	103737	4.744	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.800%	
26) 1,2-Dichloroethane-d4	5.701	65	57327	5.240	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.800%	
32) Benzene-d6	5.727	84	185328	3.822	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	76.400%	
36) 1,2-Dichloropropane-d6	6.688	67	67501	4.376	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	87.600%	
41) Toluene-d8	7.897	98	159056	3.561	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	71.200%	
43) trans-1,3-Dichloroprop...	8.180	79	16113	3.794	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	75.800%	
46) 2-Hexanone-d5	8.634	63	95954	55.645	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	111.280%	
56) 1,1,2,2-Tetrachloroeth...	10.753	84	56164	5.178	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.600%	
66) 1,2-Dichlorobenzene-d4	12.190	152	63816	4.682	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	93.600%	
Target Compounds						
5) Vinyl chloride	1.599	62	69159	6.303	ug/L	95
16) Methylene chloride	3.039	84	7312	0.589	ug/L	92
17) Methyl tert-butyl Ether	3.367	73	2750	0.129	ug/L #	85
22) cis-1,2-Dichloroethene	4.666	96	17611	1.744	ug/L	84
34) Trichloroethene	6.544	95	4079	0.351	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091323\
Data File : VU055286.D
Acq On : 13 Sep 2023 15:52
Operator : MD/SY
Sample : 04350-07DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH4M8DL

Quant Time: Sep 14 05:05:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 14 05:00:29 2023
Response via : Initial Calibration

