

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091324\
 Data File : VU060754.D
 Acq On : 13 Sep 2024 15:31
 Operator : MD/SY
 Sample : P3918-16
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AP7

Quant Time: Sep 14 00:35:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 14 00:33:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	52533	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	57900	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	25794	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	10440	4.500	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.000%
7) Chloroethane-d5	1.911	69	16777	4.772	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.400%
11) 1,1-Dichloroethene-d2	2.563	65	8653	6.328	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	126.600%#
20) 2-Butanone-d5	4.634	46	36587	45.855	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.700%
24) Chloroform-d	5.055	84	33517	5.067	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.400%
26) 1,2-Dichloroethane-d4	5.695	65	19089	5.530	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.600%
32) Benzene-d6	5.718	84	55277	4.366	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.400%
36) 1,2-Dichloropropane-d6	6.685	67	18236	4.464	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.200%
41) Toluene-d8	7.891	98	48132	4.052	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.000%
43) trans-1,3-Dichloroprop...	8.177	79	6280	4.296	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.000%
46) 2-Hexanone-d5	8.627	63	28600	40.864	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.720%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	15391	4.234	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	23255	6.114	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	122.200%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091324\
Data File : VU060754.D
Acq On : 13 Sep 2024 15:31
Operator : MD/SY
Sample : P3918-16
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AP7

Quant Time: Sep 14 00:35:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 14 00:33:41 2024
Response via : Initial Calibration

