

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091324\
 Data File : VU060761.D
 Acq On : 13 Sep 2024 18:21
 Operator : MD/SY
 Sample : P3967-02
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COAN3

Quant Time: Sep 14 00:37:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 14 00:33:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	49979	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	51891	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	23801	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	10031	4.544	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.800%
7) Chloroethane-d5	1.911	69	17733	5.302	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.000%
11) 1,1-Dichloroethene-d2	2.560	65	8268	6.355	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	127.200%#
20) 2-Butanone-d5	4.640	46	30802	40.577	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.160%
24) Chloroform-d	5.055	84	35626	5.661	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.200%
26) 1,2-Dichloroethane-d4	5.695	65	20335	6.192	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	123.800%
32) Benzene-d6	5.721	84	57173	5.039	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.800%
36) 1,2-Dichloropropane-d6	6.682	67	17951	4.903	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.000%
41) Toluene-d8	7.891	98	49114	4.614	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.200%
43) trans-1,3-Dichloroprop...	8.177	79	6796	5.187	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	103.800%
46) 2-Hexanone-d5	8.631	63	24064	38.364	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.720%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	15397	4.726	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	31861	9.079	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	181.600%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091324\
Data File : VU060761.D
Acq On : 13 Sep 2024 18:21
Operator : MD/SY
Sample : P3967-02
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
COAN3

Quant Time: Sep 14 00:37:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR082324WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 14 00:33:41 2024
Response via : Initial Calibration

