


Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091420\
Data File : VU040120.D
Acq On : 14 Sep 2020 15:14
Operator : SY/MD
Sample : L3945-18
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFSH6

Manual Integrations
APPROVED

MMDadoda
9/15/2020 8:30:18 AM

Quant Time: Sep 15 03:05:39 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 15 01:48:35 2020
Response via : Initial Calibration



Quantitation Report (Qedit)


Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091420\
Data File : VU040120.D
Acq On : 14 Sep 2020 15:14
Operator : SY/MD
Sample : L3945-18
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFSH6

Manual Integrations
APPROVED

MMDadoda
9/15/2020 8:30:18 AM

Quant Time: Sep 15 02:14:46 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 15 01:48:35 2020
Response via : Initial Calibration



(13) Acetone (T)

2.674min (-0.004) 1.91ug/L

response 4148

Ion	Exp%	Act%
43.05	100	100
58.05	27.70	31.92
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091420\
 Data File : VU040120.D
 Acq On : 14 Sep 2020 15:14
 Operator : SY/MD
 Sample : L3945-18
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFSH6

Manual Integrations
 APPROVED

MMDadoda
 9/15/2020 8:30:18 AM

Quant Time: Sep 15 02:14:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 15 01:48:35 2020
 Response via : Initial Calibration



TIC: VU040120.D

(13) Acetone (T)

2.690min (+0.013) 5.87ug/L m

response 12739

Ion	Exp%	Act%
43.05	100	100
58.05	27.70	10.39
0.00	0.00	0.00
0.00	0.00	0.00

MD
9/15/20

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091420\
 Data File : VU040120.D
 Acq On : 14 Sep 2020 15:14
 Operator : SY/MD
 Sample : L3945-18
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFSH6

Manual Integrations
 APPROVED

MMDadoda
 9/15/2020 8:30:18 AM

Quant Time: Sep 15 03:05:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 15 01:48:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	107388	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	107359	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	48892	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	27302	4.66	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.20%
7) Chloroethane-d5	1.92	69	24328	4.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.60%
11) 1,1-Dichloroethene-d2	2.58	63	46712	3.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.00%
20) 2-Butanone-d5	4.66	46	103125	43.97	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.94%
24) Chloroform-d	5.08	84	57902	4.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.60%
26) 1,2-Dichloroethane-d4	5.72	65	35835	4.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.80%
32) Benzene-d6	5.74	84	107452	4.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.80%
36) 1,2-Dichloropropane-d6	6.70	67	35005	4.24	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.80%
41) Toluene-d8	7.91	98	87332	3.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.20%
43) trans-1,3-Dichloropropene-	8.19	79	14205	3.74	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.80%
46) 2-Hexanone-d5	8.64	63	72496	40.33	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.66%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	30901	4.33	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	35152	4.44	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.80%

Target Compounds

Target Compounds				Ovalue
3) Chloromethane	1.53	50	7134	0.699 ug/L 100
13) Acetone	2.69	43	12739m	5.868 ug/L
14) Carbon disulfide	2.81	76	2682	0.101 ug/L # 93
25) Chloroform	5.11	83	6266	0.333 ug/L 95

2 MB
 09/15/20

(#) = qualifier out of range (m) = manual integration (+) = signals summed