

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091423\
 Data File : VU055320.D
 Acq On : 14 Sep 2023 22:08
 Operator : MD/SY
 Sample : VU0914WBL03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK036

Quant Time: Sep 15 06:45:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 15 02:53:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	181398	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	189143	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	89761	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	53252	6.264	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	125.200%
7) Chloroethane-d5	1.911	69	54229	5.510	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.200%
11) 1,1-Dichloroethene-d2	2.563	65	20353	4.202	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.000%
20) 2-Butanone-d5	4.634	46	103168	41.817	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.640%
24) Chloroform-d	5.062	84	95377	4.228	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.600%
26) 1,2-Dichloroethane-d4	5.698	65	49670	4.401	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.000%
32) Benzene-d6	5.724	84	193382	4.142	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.800%
36) 1,2-Dichloropropane-d6	6.689	67	60516	4.075	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.600%
41) Toluene-d8	7.897	98	167982	3.906	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.200%
43) trans-1,3-Dichloroprop...	8.177	79	17866	4.369	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.400%
46) 2-Hexanone-d5	8.631	63	60788	36.614	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	73.220%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	42045	4.026	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.600%
66) 1,2-Dichlorobenzene-d4	12.193	152	56284	4.296	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.000%
Target Compounds						Qvalue
16) Methylene chloride	3.039	84	7321	0.571	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091423\
Data File : VU055320.D
Acq On : 14 Sep 2023 22:08
Operator : MD/SY
Sample : VU0914WBL03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK036

Quant Time: Sep 15 06:45:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 15 02:53:52 2023
Response via : Initial Calibration

