

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091518\
 Data File : VU026746.D
 Acq On : 13 Sep 2018 23:19
 Operator : MD/SY
 Sample : J4893-16
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C08S5

Quant Time: Sep 14 04:03:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Sep 14 01:21:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	177547	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	163943	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	71100	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	60090	45.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.84%
7) Chloroethane-d5	1.68	69	48096	45.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.22%
11) 1,1-Dichloroethene-d2	2.27	63	82933	35.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.04%
21) 2-Butanone-d5	4.18	46	82914	100.70	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.70%
24) Chloroform-d	4.65	84	90923	41.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.66%
26) 1,2-Dichloroethane-d4	5.31	65	65730	45.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.30%
32) Benzene-d6	5.34	84	202703	47.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.54%
36) 1,2-Dichloropropane-d6	6.33	67	68804	48.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.64%
41) Toluene-d8	7.57	98	178923	44.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.88%
43) trans-1,3-Dichloropropene-	7.85	79	28457	42.99	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.98%
47) 2-Hexanone-d5	8.31	63	58321	100.39	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.39%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	88684	45.49	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.98%
64) 1,2-Dichlorobenzene-d4	11.86	152	68324	48.20	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.40%

Target Compounds

					Ovalue
13) Acetone	2.32	43	2545	2.714 ug/L	93
51) Chlorobenzene	9.12	112	9235	2.956 ug/L	96
63) 1,4-Dichlorobenzene	11.51	146	9331	4.315 ug/L	99
65) 1,2-Dichlorobenzene	11.88	146	12365	5.731 ug/L	98
68) 1,2,4-trichlorobenzene	13.51	180	4619	3.282 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091518\
Data File : VU026746.D
Acq On : 13 Sep 2018 23:19
Operator : MD/SY
Sample : J4893-16
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C08S5

Quant Time: Sep 14 04:03:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091018WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 14 01:21:50 2018
Response via : Initial Calibration

