

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091518\
 Data File : VU026764.D
 Acq On : 14 Sep 2018 14:26
 Operator : MD/SY
 Sample : J4893-20DL 200X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C08K8DL

Quant Time: Sep 15 01:49:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091018WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 15 00:18:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	160608	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	150447	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	64641	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	53228	44.47	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.94%
7) Chloroethane-d5	1.68	69	44719	46.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.74%
11) 1,1-Dichloroethene-d2	2.27	63	74397	34.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.46%
21) 2-Butanone-d5	4.18	46	79182	106.31	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.31%
24) Chloroform-d	4.65	84	92195	46.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.66%
26) 1,2-Dichloroethane-d4	5.31	65	62596	48.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.12%
32) Benzene-d6	5.34	84	185645	47.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.36%
36) 1,2-Dichloropropane-d6	6.33	67	64658	49.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.96%
41) Toluene-d8	7.57	98	161509	44.21	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.42%
43) trans-1,3-Dichloropropene-	7.85	79	27248	44.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.70%
47) 2-Hexanone-d5	8.31	63	54088	101.45	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.45%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	85986	48.06	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.12%
64) 1,2-Dichlorobenzene-d4	11.86	152	63669	49.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.82%

Target Compounds

					Ovalue
33) Benzene	5.39	78	41270	9.796	ug/L 100
51) Chlorobenzene	9.12	112	63541	22.161	ug/L 98
62) 1,3-Dichlorobenzene	11.42	146	5950	3.206	ug/L 96
63) 1,4-Dichlorobenzene	11.51	146	64727	32.919	ug/L 99
65) 1,2-Dichlorobenzene	11.88	146	151950	77.462	ug/L 100
68) 1,2,4-trichlorobenzene	13.51	180	7517	5.874	ug/L 98
70) 1,2,3-Trichlorobenzene	14.00	180	1939	1.399	ug/L # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091518\
Data File : VU026764.D
Acq On : 14 Sep 2018 14:26
Operator : MD/SY
Sample : J4893-20DL 200X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C08K8DL

Quant Time: Sep 15 01:49:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091018WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 15 00:18:06 2018
Response via : Initial Calibration

