

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091618\
 Data File : VU026794.D
 Acq On : 15 Sep 2018 22:20
 Operator : MD/SY
 Sample : J4918-01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0KX1

Quant Time: Sep 17 08:18:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR091618WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Sep 17 07:09:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	119339	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	99716	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	40733	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	43417	5.06	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.20%
7) Chloroethane-d5	1.68	69	36520	5.16	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.20%
11) 1,1-Dichloroethene-d2	2.28	63	70529	4.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.00%
20) 2-Butanone-d5	4.20	46	97385	48.68	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.36%
24) Chloroform-d	4.65	84	74221	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.60%
26) 1,2-Dichloroethane-d4	5.31	65	39859	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.80%
32) Benzene-d6	5.34	84	152723	5.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.60%
36) 1,2-Dichloropropane-d6	6.33	67	49215	5.49	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.80%
41) Toluene-d8	7.57	98	114907	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.00%
43) trans-1,3-Dichloropropene-	7.85	79	14164	4.26	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.20%
46) 2-Hexanone-d5	8.32	63	58170	44.30	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.60%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	35242	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	42008	5.56	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.20%

Target Compounds						Ovalue
12) 1,1-Dichloroethene	2.29	96	18169	2.04	ug/L	# 63
13) Acetone	2.33	43	4803	2.70	ug/L	91
34) Trichloroethene	6.18	95	285380	30.00	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091618\
Data File : VU026794.D
Acq On : 15 Sep 2018 22:20
Operator : MD/SY
Sample : J4918-01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0KX1

Quant Time: Sep 17 08:18:22 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR091618WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 17 07:09:54 2018
Response via : Initial Calibration

