

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091618\
 Data File : VU026796.D
 Acq On : 15 Sep 2018 23:12
 Operator : MD/SY
 Sample : J4918-03
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0KX3

Quant Time: Sep 17 08:14:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR091618WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Sep 17 07:09:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	117627	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	95860	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	37847	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	43964	5.20	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.00%
7) Chloroethane-d5	1.68	69	37220	5.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.80%
11) 1,1-Dichloroethene-d2	2.28	63	77012	4.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	94.20%
20) 2-Butanone-d5	4.20	46	96111	48.74	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.48%
24) Chloroform-d	4.65	84	74377	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
26) 1,2-Dichloroethane-d4	5.31	65	39641	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.60%
32) Benzene-d6	5.34	84	154309	5.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.00%
36) 1,2-Dichloropropane-d6	6.33	67	49220	5.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	114.20%
41) Toluene-d8	7.57	98	115776	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.20%
43) trans-1,3-Dichloropropene-	7.85	79	14366	4.49	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.80%
46) 2-Hexanone-d5	8.32	63	56464	44.73	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.46%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	35226	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	40063	5.70	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	114.00%

Target Compounds						Ovalue
12) 1,1-Dichloroethene	2.29	96	28906	3.30	ug/L	# 79
13) Acetone	2.33	43	5655	3.23	ug/L	94
22) cis-1,2-Dichloroethene	4.23	96	15942	1.53	ug/L	89
29) 1,1,1-Trichloroethane	4.91	97	6400	0.46	ug/L	99
34) Trichloroethene	6.18	95	455371	49.80	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091618\
Data File : VU026796.D
Acq On : 15 Sep 2018 23:12
Operator : MD/SY
Sample : J4918-03
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0KX3

Quant Time: Sep 17 08:14:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR091618WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 17 07:09:54 2018
Response via : Initial Calibration

