

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091618\
 Data File : VU026806.D
 Acq On : 16 Sep 2018 04:22
 Operator : MD/SY
 Sample : J4918-13
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0KY3

Quant Time: Sep 17 07:48:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR091618WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Sep 17 07:09:54 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	92322	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	70084	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	27112	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	35369	5.33	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.60%
7) Chloroethane-d5	1.68	69	33595	6.14	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	122.80%
11) 1,1-Dichloroethene-d2	2.28	63	45323	3.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.60%
20) 2-Butanone-d5	4.20	46	102143	66.00	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	132.00%#
24) Chloroform-d	4.65	84	66958	5.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	116.00%
26) 1,2-Dichloroethane-d4	5.31	65	38637	6.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	127.40%
32) Benzene-d6	5.34	84	131639	6.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	134.40%#
36) 1,2-Dichloropropane-d6	6.33	67	44510	7.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	141.20%#
41) Toluene-d8	7.57	98	85716	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
43) trans-1,3-Dichloropropene-	7.85	79	12745	5.45	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	109.00%
46) 2-Hexanone-d5	8.32	63	59661	64.64	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	129.28%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	35344	7.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	143.40%#
64) 1,2-Dichlorobenzene-d4	11.86	152	33083	6.57	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	131.40%#

Target Compounds					Ovalue
13) Acetone	2.33	43	2191	1.59 ug/L	91
34) Trichloroethene	6.19	95	4334	0.65 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091618\
Data File : VU026806.D
Acq On : 16 Sep 2018 04:22
Operator : MD/SY
Sample : J4918-13
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0KY3

Quant Time: Sep 17 07:48:07 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR091618WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Sep 17 07:09:54 2018
Response via : Initial Calibration

