

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091618\  
 Data File : VU026814.D  
 Acq On : 17 Sep 2018 10:17  
 Operator : MD/SY  
 Sample : VU0916WBL03  
 Misc : 25.0mL/MSVOA U/WATER  
 ALS Vial : 47 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VBLK60

Quant Time: Sep 17 15:58:47 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR091618WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Mon Sep 17 07:09:54 2018  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.89  | 114  | 133087   | 5.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 9.09  | 117  | 107784   | 5.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.48 | 152  | 43424    | 5.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |         |
|--------------------------------|--------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 1.40   | 65    | 43085    | 4.51     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 40 - 130 | Recovery | =    | 90.20%  |
| 7) Chloroethane-d5             | 1.68   | 69    | 37862    | 4.80     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 130 | Recovery | =    | 96.00%  |
| 11) 1,1-Dichloroethene-d2      | 2.27   | 63    | 62165    | 3.36     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 125 | Recovery | =    | 67.20%  |
| 20) 2-Butanone-d5              | 4.21   | 46    | 107207   | 48.05    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 40 - 130 | Recovery | =    | 96.10%  |
| 24) Chloroform-d               | 4.65   | 84    | 79935    | 4.81     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 96.20%  |
| 26) 1,2-Dichloroethane-d4      | 5.31   | 65    | 41904    | 4.80     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 96.00%  |
| 32) Benzene-d6                 | 5.34   | 84    | 162809   | 5.40     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 108.00% |
| 36) 1,2-Dichloropropane-d6     | 6.33   | 67    | 51493    | 5.31     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 140 | Recovery | =    | 106.20% |
| 41) Toluene-d8                 | 7.57   | 98    | 126058   | 4.52     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 90.40%  |
| 43) trans-1,3-Dichloropropene- | 7.85   | 79    | 17196    | 4.78     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 55 - 130 | Recovery | =    | 95.60%  |
| 46) 2-Hexanone-d5              | 8.33   | 63    | 67228    | 47.36    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 45 - 130 | Recovery | =    | 94.72%  |
| 57) 1,1,2,2-Tetrachloroethane- | 10.43  | 84    | 37389    | 4.93     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 120 | Recovery | =    | 98.60%  |
| 64) 1,2-Dichlorobenzene-d4     | 11.86  | 152   | 44533    | 5.53     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 80 - 120 | Recovery | =    | 110.60% |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091618\  
Data File : VU026814.D  
Acq On : 17 Sep 2018 10:17  
Operator : MD/SY  
Sample : VU0916WBL03  
Misc : 25.0mL/MSVOA U/WATER  
ALS Vial : 47 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VBLK60

Quant Time: Sep 17 15:58:47 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR091618WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Mon Sep 17 07:09:54 2018  
Response via : Initial Calibration

