

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091620\
 Data File : VU040143.D
 Acq On : 16 Sep 2020 11:12
 Operator : SY/MD
 Sample : VU0916WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK07

Quant Time: Sep 17 01:17:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 17 01:09:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	112450	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	112504	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	51983	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	28923	4.71	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.20%
7) Chloroethane-d5	1.92	69	25478	4.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.60%
11) 1,1-Dichloroethene-d2	2.58	63	47792	3.03	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.60%
20) 2-Butanone-d5	4.67	46	107494	43.77	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.54%
24) Chloroform-d	5.08	84	57002	3.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.40%
26) 1,2-Dichloroethane-d4	5.72	65	38845	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.80%
32) Benzene-d6	5.75	84	110390	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.00%
36) 1,2-Dichloropropane-d6	6.70	67	36099	4.17	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.40%
41) Toluene-d8	7.91	98	88602	3.73	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.60%
43) trans-1,3-Dichloropropene-	8.19	79	15154	3.81	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.20%
46) 2-Hexanone-d5	8.64	63	75940	40.31	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.62%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	32047	4.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	34763	4.13	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	82.60%

Target Compounds					Ovalue
16) Methylene chloride	3.07	84	1521	0.133 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091620\
Data File : VU040143.D
Acq On : 16 Sep 2020 11:12
Operator : SY/MD
Sample : VU0916WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK07

Quant Time: Sep 17 01:17:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 17 01:09:36 2020
Response via : Initial Calibration

