

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091620\
 Data File : VU040146.D
 Acq On : 16 Sep 2020 14:30
 Operator : SY/MD
 Sample : L4041-02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AA6

Quant Time: Sep 17 01:26:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 17 01:09:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	109630	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	110040	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	46716	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	26423	4.41	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.20%
7) Chloroethane-d5	1.93	69	24955	5.00	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.00%
11) 1,1-Dichloroethene-d2	2.58	63	45908	2.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.80%#
20) 2-Butanone-d5	4.66	46	101287	42.31	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.62%
24) Chloroform-d	5.08	84	57625	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.40%
26) 1,2-Dichloroethane-d4	5.72	65	37756	4.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.40%
32) Benzene-d6	5.74	84	107941	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.00%
36) 1,2-Dichloropropane-d6	6.70	67	36719	4.33	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.60%
41) Toluene-d8	7.91	98	87249	3.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.20%
43) trans-1,3-Dichloropropene-	8.19	79	14198	3.65	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.00%
46) 2-Hexanone-d5	8.64	63	70674	38.35	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.70%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	31929	4.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	33665	4.45	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091620\
Data File : VU040146.D
Acq On : 16 Sep 2020 14:30
Operator : SY/MD
Sample : L4041-02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AA6

Quant Time: Sep 17 01:26:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 17 01:09:36 2020
Response via : Initial Calibration

