

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091620\
 Data File : VU040150.D
 Acq On : 16 Sep 2020 16:03
 Operator : SY/MD
 Sample : L4041-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AB2

Quant Time: Sep 17 01:37:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Sep 17 01:09:36 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	121844	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	122720	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	51605	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	27711	4.17	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.40%
7) Chloroethane-d5	1.92	69	25300	4.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.20%
11) 1,1-Dichloroethene-d2	2.58	63	46033	2.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.00%#
20) 2-Butanone-d5	4.66	46	107975	40.58	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.16%
24) Chloroform-d	5.09	84	61925	3.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.60%
26) 1,2-Dichloroethane-d4	5.72	65	36918	3.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.00%
32) Benzene-d6	5.75	84	111206	3.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.60%
36) 1,2-Dichloropropane-d6	6.70	67	37112	3.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.60%
41) Toluene-d8	7.91	98	86677	3.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	67.00%#
43) trans-1,3-Dichloropropene-	8.18	79	14531	3.35	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.00%
46) 2-Hexanone-d5	8.64	63	74733	36.37	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.74%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	32267	3.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	34448	4.12	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	82.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091620\
Data File : VU040150.D
Acq On : 16 Sep 2020 16:03
Operator : SY/MD
Sample : L4041-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AB2

Quant Time: Sep 17 01:37:17 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Sep 17 01:09:36 2020
Response via : Initial Calibration

