

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091622\
 Data File : VU050725.D
 Acq On : 16 Sep 2022 15:02
 Operator : JC/MD
 Sample : N4568-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 38025-110121

Quant Time: Sep 17 01:07:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081922DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Aug 25 14:01:59 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.115	96	44941	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.636	95	12446	0.813	ug/l	0.00
Spiked Amount 1.000			Recovery	=	81.000%	
68) 1,2-Dichlorobenzene-d4	12.198	152	14620	0.808	ug/l	0.00
Spiked Amount 1.000			Recovery	=	81.000%	
Target Compounds						
						Qvalue
15) Methylene Chloride	3.044	84	21018	1.174	ug/l	97
27) Chloroform	5.089	83	920843	33.351	ug/l	98
43) Dibromomethane	6.919	93	4511	0.562	ug/l	93
44) Bromodichloromethane	7.105	83	1040801	54.948	ug/l	97
55) Dibromochloromethane	8.809	129	219477	16.912	ug/l	98
67) Bromoform	10.291	173	201908	27.396	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091622\
Data File : VU050725.D
Acq On : 16 Sep 2022 15:02
Operator : JC/MD
Sample : N4568-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
38025-110121

Quant Time: Sep 17 01:07:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U081922DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Aug 25 14:01:59 2022
Response via : Initial Calibration

