

Data File : VU026824.D
Acq On : 17 Sep 2018 15:02
Operator : MD/SY
Sample : VIBLK40
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK40

Quant Time: Sep 18 01:30:33 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091718WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 18 00:50:30 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	283110	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	267946	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	116871	50.00	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.40	65	81615	38.97	ug/L	0.00
7) Chloroethane-d5	1.68	69	70903	39.55	ug/L	0.00
11) 1,1-Dichloroethene-d2	2.27	63	116289	29.35	ug/L	0.00
21) 2-Butanone-d5	4.18	46	106766	76.82	ug/L	0.00
24) Chloroform-d	4.65	84	147248	37.90	ug/L	0.00
26) 1,2-Dichloroethane-d4	5.31	65	95099	38.88	ug/L	0.00
32) Benzene-d6	5.34	84	303730	40.45	ug/L	0.00
36) 1,2-Dichloropropane-d6	6.33	67	97343	40.23	ug/L	0.00
41) Toluene-d8	7.57	98	262083	38.63	ug/L	0.00
43) trans-1,3-Dichloropropene-	7.85	79	42104	38.98	ug/L	0.00
47) 2-Hexanone-d5	8.31	63	72945	76.71	ug/L	0.00
57) 1,1,2,2-Tetrachloroethane-	10.43	84	126080	37.37	ug/L	0.00
64) 1,2-Dichlorobenzene-d4	11.86	152	102979	42.36	ug/L	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU026824.D
Acq On : 17 Sep 2018 15:02
Operator : MD/SY
Sample : VIBLK40
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK40

Quant Time: Sep 18 01:30:33 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091718WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Sep 18 00:50:30 2018
Response via : Initial Calibration

