

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091719\
 Data File : VU034609.D
 Acq On : 17 Sep 2019 12:41
 Operator : JC/SP
 Sample : K4894-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COMRO

Quant Time: Sep 18 05:22:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091619WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 18 04:30:59 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	234299	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	219351	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.46	152	85528	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	52586	46.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.04%
7) Chloroethane-d5	1.67	69	46740	53.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.52%
11) 1,1-Dichloroethene-d2	2.26	63	84799	36.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.32%
21) 2-Butanone-d5	4.15	46	152260	119.46	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	119.46%
24) Chloroform-d	4.62	84	137937	53.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.50%
26) 1,2-Dichloroethane-d4	5.29	65	90839	58.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	116.02%
32) Benzene-d6	5.32	84	253782	52.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.24%
36) 1,2-Dichloropropane-d6	6.31	67	98354	54.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	109.56%
41) Toluene-d8	7.54	98	238598	50.95	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.90%
43) trans-1,3-Dichloropropene-	7.83	79	44202	51.61	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	103.22%
47) 2-Hexanone-d5	8.29	63	106206	109.05	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.05%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	149490	53.29	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.58%
64) 1,2-Dichlorobenzene-d4	11.84	152	82859	54.31	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.62%


Target Compounds					Ovalue
13) Acetone	2.31	43	7565	6.604 ug/L	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091719\
Data File : VU034609.D
Acq On : 17 Sep 2019 12:41
Operator : JC/SP
Sample : K4894-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
COMRO

Quant Time: Sep 18 05:22:15 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091619WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 18 04:30:59 2019
Response via : Initial Calibration





#13

Acetone

Concen: 6.604 ug/L

RT: 2.31 min Scan# 378

Delta R.T. 0.00 min

Lab File: VU034609.D

Acq: 17 Sep 2019 12:41

Instrument :
MSVOA_U
ClientSampleId :
COMRO

Tot Ion: 43 Resp: 7565

Ion Ratio Lower Upper

43 100

58 35.0 0.0 65.0

