

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091720\
 Data File : VU040196.D
 Acq On : 17 Sep 2020 21:03
 Operator : SY/MD
 Sample : L4063-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AJ2

Quant Time: Sep 18 05:07:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 18 04:01:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	101463	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	100938	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	41575	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	27039	4.88	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.60%
7) Chloroethane-d5	1.92	69	25218	5.46	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.20%
11) 1,1-Dichloroethene-d2	2.58	63	46142	3.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.00%
20) 2-Butanone-d5	4.66	46	114854	51.83	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.66%
24) Chloroform-d	5.08	84	59759	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.40%
26) 1,2-Dichloroethane-d4	5.72	65	40244	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.80%
32) Benzene-d6	5.74	84	117716	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.80%
36) 1,2-Dichloropropane-d6	6.70	67	40044	5.15	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.00%
41) Toluene-d8	7.91	98	96130	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.40%
43) trans-1,3-Dichloropropene-	8.19	79	15623	4.38	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.60%
46) 2-Hexanone-d5	8.64	63	82015	48.52	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.04%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	33761	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	37132	5.52	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.40%

Target Compounds					Ovalue
25) Chloroform	5.11	83	8251	0.463 ug/L	95
34) Trichloroethene	6.55	95	1065	0.110 ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091720\
Data File : VU040196.D
Acq On : 17 Sep 2020 21:03
Operator : SY/MD
Sample : L4063-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AJ2

Quant Time: Sep 18 05:07:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 18 04:01:22 2020
Response via : Initial Calibration

