

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091721\
 Data File : VU044680.D
 Acq On : 17 Sep 2021 12:40
 Operator : SY/MD
 Sample : M3763-17
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Manual Integrations
 APPROVED

SAM
 9/18/2021 11:23:53 AM

Quant Time: Sep 18 01:33:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090921WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 18 01:28:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	124678	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	124847	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	53090	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	24630	3.348	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	67.000%	
7) Chloroethane-d5	1.919	69	28928	3.853	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	77.000%	
11) 1,1-Dichloroethene-d2	2.571	65	13714	3.691	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	73.800%	
20) 2-Butanone-d5	4.645	46	162204m	52.599	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	105.200%	
24) Chloroform-d	5.070	84	70303	4.543	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.800%	
26) 1,2-Dichloroethane-d4	5.710	65	42903	4.650	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.000%	
32) Benzene-d6	5.735	84	129389	4.065	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.200%	
36) 1,2-Dichloropropane-d6	6.700	67	48133	4.604	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	92.000%	
41) Toluene-d8	7.902	98	102253	3.690	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	73.800%	
43) trans-1,3-Dichloroprop...	8.185	79	14204	3.854	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	77.000%	
46) 2-Hexanone-d5	8.645	63	116498	49.186	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	98.380%	
56) 1,1,2,2-Tetrachloroeth...	10.761	84	43380	4.680	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	93.600%	
66) 1,2-Dichlorobenzene-d4	12.198	152	38468	4.391	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	87.800%	
Target Compounds						Qvalue
16) Methylene chloride	3.057	84	9750	0.765	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091721\
Data File : VU044680.D
Acq On : 17 Sep 2021 12:40
Operator : SY/MD
Sample : M3763-17
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Manual Integrations
APPROVED

SAM
9/18/2021 11:23:53 AM

Quant Time: Sep 18 01:33:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090921WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 18 01:28:45 2021
Response via : Initial Calibration

