

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU091721\  
 Data File : VU044682.D  
 Acq On : 17 Sep 2021 13:28  
 Operator : SY/MD  
 Sample : M3739-18  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 C0AG0

Quant Time: Sep 18 01:33:22 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR090921WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Sat Sep 18 01:28:45 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon           | Response | Conc   | Units    | Dev(Min) |
|-------------------------------|--------|----------------|----------|--------|----------|----------|
| -----                         |        |                |          |        |          |          |
| Internal Standards            |        |                |          |        |          |          |
| 1) 1,4-Difluorobenzene        | 6.256  | 114            | 121552   | 5.000  | ug/L     | 0.00     |
| 28) Chlorobenzene-d5          | 9.423  | 117            | 123763   | 5.000  | ug/L     | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.816 | 152            | 53591    | 5.000  | ug/L     | 0.00     |
|                               |        |                |          |        |          |          |
| System Monitoring Compounds   |        |                |          |        |          |          |
| 4) Vinyl Chloride-d3          | 1.601  | 65             | 23825    | 3.322  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 40 - 130 | Recovery | =      | 66.400%  |          |
| 7) Chloroethane-d5            | 1.919  | 69             | 28875    | 3.945  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 130 | Recovery | =      | 79.000%  |          |
| 11) 1,1-Dichloroethene-d2     | 2.572  | 65             | 12519    | 3.456  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 125 | Recovery | =      | 69.200%  |          |
| 20) 2-Butanone-d5             | 4.642  | 46             | 157964   | 52.541 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 40 - 130 | Recovery | =      | 105.080% |          |
| 24) Chloroform-d              | 5.070  | 84             | 69874    | 4.632  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 92.600%  |          |
| 26) 1,2-Dichloroethane-d4     | 5.710  | 65             | 42387    | 4.712  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 94.200%  |          |
| 32) Benzene-d6                | 5.736  | 84             | 129308   | 4.098  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 82.000%  |          |
| 36) 1,2-Dichloropropane-d6    | 6.700  | 67             | 46829    | 4.519  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 140 | Recovery | =      | 90.400%  |          |
| 41) Toluene-d8                | 7.903  | 98             | 98749    | 3.595  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 71.800%  |          |
| 43) trans-1,3-Dichloroprop... | 8.186  | 79             | 13607    | 3.725  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 55 - 130 | Recovery | =      | 74.400%  |          |
| 46) 2-Hexanone-d5             | 8.645  | 63             | 115280   | 49.099 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 45 - 130 | Recovery | =      | 98.200%  |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.761 | 84             | 42326    | 4.606  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 120 | Recovery | =      | 92.200%  |          |
| 66) 1,2-Dichlorobenzene-d4    | 12.198 | 152            | 38422    | 4.345  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 80 - 120 | Recovery | =      | 86.800%  |          |
|                               |        |                |          |        |          |          |
| Target Compounds              |        |                |          |        | Qvalue   |          |
| 13) Acetone                   | 2.649  | 43             | 8634     | 4.409  | ug/L     | 81       |
| -----                         |        |                |          |        |          |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU091721\  
Data File : VU044682.D  
Acq On : 17 Sep 2021 13:28  
Operator : SY/MD  
Sample : M3739-18  
Misc : 25.0mL/MSVOA\_U/WATER  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
C0AG0

Quant Time: Sep 18 01:33:22 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR090921WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Sat Sep 18 01:28:45 2021  
Response via : Initial Calibration

