

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091721\
 Data File : VU044689.D
 Acq On : 17 Sep 2021 16:42
 Operator : SY/MD
 Sample : M3790-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CH2

Quant Time: Sep 18 01:34:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090921WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 18 01:28:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	124834	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	123929	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	53821	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	21972	2.983	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	59.600%
7) Chloroethane-d5	1.919	69	26161	3.480	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	69.600%
11) 1,1-Dichloroethene-d2	2.572	65	12182	3.274	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.400%
20) 2-Butanone-d5	4.645	46	154878	50.160	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.320%
24) Chloroform-d	5.073	84	64987	4.195	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.800%
26) 1,2-Dichloroethane-d4	5.710	65	39368	4.261	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.200%
32) Benzene-d6	5.735	84	119006	3.766	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.400%
36) 1,2-Dichloropropane-d6	6.697	67	43044	4.148	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.000%
41) Toluene-d8	7.906	98	92593	3.366	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	67.400%#
43) trans-1,3-Dichloroprop...	8.186	79	13297	3.635	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.600%
46) 2-Hexanone-d5	8.645	63	114413	48.664	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.320%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	40247	4.374	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.400%
66) 1,2-Dichlorobenzene-d4	12.198	152	36488	4.109	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	82.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091721\
Data File : VU044689.D
Acq On : 17 Sep 2021 16:42
Operator : SY/MD
Sample : M3790-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CH2

Quant Time: Sep 18 01:34:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090921WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 18 01:28:45 2021
Response via : Initial Calibration

