

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091721\
 Data File : VU044692.D
 Acq On : 17 Sep 2021 17:55
 Operator : SY/MD
 Sample : M3790-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0CJ1

Manual Integrations
 APPROVED

SAM
 9/18/2021 11:24:25 AM

Quant Time: Sep 18 01:35:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090921WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 18 01:28:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	117423	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	116982	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	49554	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	21249	3.067	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.400%
7) Chloroethane-d5	1.919	69	26170	3.701	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	74.000%
11) 1,1-Dichloroethene-d2	2.572	65	12514	3.576	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.600%
20) 2-Butanone-d5	4.646	46	158203m	54.471	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.940%
24) Chloroform-d	5.070	84	65742	4.511	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.200%
26) 1,2-Dichloroethane-d4	5.710	65	39818	4.582	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.600%
32) Benzene-d6	5.735	84	118867	3.985	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.800%
36) 1,2-Dichloropropane-d6	6.697	67	44585	4.551	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.000%
41) Toluene-d8	7.903	98	90237	3.475	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.600%#
43) trans-1,3-Dichloroprop...	8.186	79	12650	3.663	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.200%
46) 2-Hexanone-d5	8.645	63	113991	51.364	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.720%
56) 1,1,2,2-Tetrachloroeth...	10.761	84	41436	4.771	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.400%
66) 1,2-Dichlorobenzene-d4	12.198	152	35490	4.340	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091721\
Data File : VU044692.D
Acq On : 17 Sep 2021 17:55
Operator : SY/MD
Sample : M3790-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0CJ1

Manual Integrations
APPROVED

SAM
9/18/2021 11:24:25 AM

Quant Time: Sep 18 01:35:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090921WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 18 01:28:45 2021
Response via : Initial Calibration

