

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091721\
 Data File : VU044687.D
 Acq On : 17 Sep 2021 15:30
 Operator : SY/MD
 Sample : M3777-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B78

Manual Integrations
 APPROVED

SAM
 9/18/2021 11:24:05 AM

Quant Time: Sep 18 01:34:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090921WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Sep 18 01:28:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	120436	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	121423	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	52235	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	23427	3.297	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	66.000%	
7) Chloroethane-d5	1.919	69	28062	3.870	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	77.400%	
11) 1,1-Dichloroethene-d2	2.571	65	13141	3.661	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	73.200%	
20) 2-Butanone-d5	4.645	46	163949m	55.037	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	110.080%	
24) Chloroform-d	5.070	84	68495	4.582	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.600%	
26) 1,2-Dichloroethane-d4	5.710	65	42708	4.792	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	95.800%	
32) Benzene-d6	5.735	84	126377	4.082	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.600%	
36) 1,2-Dichloropropane-d6	6.697	67	47392	4.661	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.200%	
41) Toluene-d8	7.902	98	98730	3.663	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	73.200%	
43) trans-1,3-Dichloroprop...	8.185	79	13408	3.741	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	74.800%	
46) 2-Hexanone-d5	8.645	63	116395	50.529	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	101.060%	
56) 1,1,2,2-Tetrachloroeth...	10.761	84	43247	4.797	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	96.000%	
66) 1,2-Dichlorobenzene-d4	12.198	152	37301	4.328	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	86.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091721\
Data File : VU044687.D
Acq On : 17 Sep 2021 15:30
Operator : SY/MD
Sample : M3777-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B78

Manual Integrations
APPROVED

SAM
9/18/2021 11:24:05 AM

Quant Time: Sep 18 01:34:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR090921WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Sep 18 01:28:45 2021
Response via : Initial Calibration

