

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091724\
 Data File : VU060831.D
 Acq On : 17 Sep 2024 14:29
 Operator : MD/SY
 Sample : P3967-18
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AP9

Quant Time: Sep 18 01:18:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 18 01:15:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	143154	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	152481	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	61236	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	49658	3.857	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.200%
7) Chloroethane-d5	1.911	69	37991	4.144	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.800%
11) 1,1-Dichloroethene-d2	2.560	65	19513	4.276	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.600%
20) 2-Butanone-d5	4.637	46	96481	46.676	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.360%
24) Chloroform-d	5.055	84	87041	4.422	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
26) 1,2-Dichloroethane-d4	5.698	65	43250	4.597	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.000%
32) Benzene-d6	5.721	84	169499	4.147	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.000%
36) 1,2-Dichloropropane-d6	6.682	67	55288	4.372	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.400%
41) Toluene-d8	7.891	98	142893	3.835	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.600%
43) trans-1,3-Dichloroprop...	8.177	79	19321	4.245	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.800%
46) 2-Hexanone-d5	8.631	63	72472	40.364	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.720%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	43897	4.431	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	49710	4.831	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091724\
Data File : VU060831.D
Acq On : 17 Sep 2024 14:29
Operator : MD/SY
Sample : P3967-18
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AP9

Quant Time: Sep 18 01:18:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 18 01:15:55 2024
Response via : Initial Calibration

