

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091725\
 Data File : VU063603.D
 Acq On : 17 Sep 2025 13:30
 Operator : MD/SY
 Sample : VU0917WBL01
 Misc : 28mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK166

Quant Time: Sep 18 06:26:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 18 06:17:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.238	114	159097	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	145963	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	65433	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	34957	4.274	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	85.400%	
7) Chloroethane-d5	1.901	69	34421	4.378	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	87.600%	
11) 1,1-Dichloroethene-d2	2.554	65	16526	4.235	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	84.800%	
20) 2-Butanone-d5	4.618	46	98741	41.952	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	83.900%	
24) Chloroform-d	5.049	84	81080	4.148	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	83.000%	
26) 1,2-Dichloroethane-d4	5.692	65	45867	4.400	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.000%	
32) Benzene-d6	5.714	84	159106	4.539	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	90.800%	
36) 1,2-Dichloropropane-d6	6.679	67	52772	4.652	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.000%	
41) Toluene-d8	7.888	98	138885	4.416	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.400%	
43) trans-1,3-Dichloroprop...	8.174	79	12368	3.982	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	79.600%	
46) 2-Hexanone-d5	8.624	63	72221	44.142	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	88.280%	
56) 1,1,2,2-Tetrachloroeth...	10.746	84	38953	4.259	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	85.200%	
66) 1,2-Dichlorobenzene-d4	12.187	152	48213	4.742	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	94.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091725\
Data File : VU063603.D
Acq On : 17 Sep 2025 13:30
Operator : MD/SY
Sample : VU0917WBL01
Misc : 28mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK166

Quant Time: Sep 18 06:26:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 18 06:17:12 2025
Response via : Initial Calibration

