

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091818\
 Data File : VU026849.D
 Acq On : 18 Sep 2018 10:12
 Operator : MD/SY
 Sample : J4896-19
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C08T6

Quant Time: Sep 19 05:52:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091718WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 19 05:05:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	243658	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	229140	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	95541	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	81258	45.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.16%
7) Chloroethane-d5	1.68	69	74007	47.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.94%
11) 1,1-Dichloroethene-d2	2.27	63	118729	34.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.62%
21) 2-Butanone-d5	4.18	46	130130	108.79	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.79%
24) Chloroform-d	4.65	84	160029	47.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.72%
26) 1,2-Dichloroethane-d4	5.31	65	105256	50.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.00%
32) Benzene-d6	5.34	84	326051	50.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.54%
36) 1,2-Dichloropropane-d6	6.33	67	106613	51.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.06%
41) Toluene-d8	7.57	98	277467	47.83	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.66%
43) trans-1,3-Dichloropropene-	7.85	79	44245	47.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.78%
47) 2-Hexanone-d5	8.31	63	90100	110.80	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.80%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	144088	49.94	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.88%
64) 1,2-Dichlorobenzene-d4	11.86	152	108538	54.61	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.22%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091818\
Data File : VU026849.D
Acq On : 18 Sep 2018 10:12
Operator : MD/SY
Sample : J4896-19
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C08T6

Quant Time: Sep 19 05:52:30 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091718WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 19 05:05:35 2018
Response via : Initial Calibration

