

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091818\
 Data File : VU026872.D
 Acq On : 18 Sep 2018 19:07
 Operator : MD/SY
 Sample : VU0918MBL02
 Misc : 5.00µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK44

Quant Time: Sep 19 07:44:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091718WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 19 07:38:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	244924	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	232604	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	95912	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	87371	48.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.44%
7) Chloroethane-d5	1.68	69	77130	49.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.48%
11) 1,1-Dichloroethene-d2	2.27	63	125775	36.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.38%
21) 2-Butanone-d5	4.18	46	133435	110.98	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.98%
24) Chloroform-d	4.65	84	160534	47.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.54%
26) 1,2-Dichloroethane-d4	5.31	65	107020	50.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.14%
32) Benzene-d6	5.34	84	332732	51.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.08%
36) 1,2-Dichloropropane-d6	6.33	67	108913	51.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.70%
41) Toluene-d8	7.57	98	284052	48.23	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.46%
43) trans-1,3-Dichloropropene-	7.85	79	42823	45.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.32%
47) 2-Hexanone-d5	8.31	63	93116	112.80	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.80%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	145640	49.73	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.46%
64) 1,2-Dichlorobenzene-d4	11.86	152	109667	54.96	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.92%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091818\
Data File : VU026872.D
Acq On : 18 Sep 2018 19:07
Operator : MD/SY
Sample : VU0918MBL02
Misc : 5.00µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK44

Quant Time: Sep 19 07:44:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091718WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 19 07:38:56 2018
Response via : Initial Calibration

