

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091818\
 Data File : VU026886.D
 Acq On : 19 Sep 2018 00:32
 Operator : MD/SY
 Sample : J4865-20ME
 Misc : 4.59g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DB3P8ME

Quant Time: Sep 19 08:17:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091718WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Sep 19 07:38:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	230287	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.10	117	228671	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	129110	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	50923	29.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	59.78%#
7) Chloroethane-d5	1.64	69	32142	22.04	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	44.08%#
11) 1,1-Dichloroethene-d2	2.23	63	79632	24.70	ug/L	-0.04
Spiked Amount	50.000	Range	60 - 125	Recovery	=	49.40%#
21) 2-Butanone-d5	4.20	46	115813	102.45	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.45%
24) Chloroform-d	4.65	84	125637	39.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.52%
26) 1,2-Dichloroethane-d4	5.31	65	84087	42.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.52%
32) Benzene-d6	5.33	84	243093	37.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.86%
36) 1,2-Dichloropropane-d6	6.33	67	85073	41.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	82.40%
41) Toluene-d8	7.57	98	200645	34.66	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	69.32%#
43) trans-1,3-Dichloropropene-	7.85	79	34210	37.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.22%
47) 2-Hexanone-d5	8.33	63	85614	105.50	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.50%
57) 1,1,2,2-Tetrachloroethane-	10.44	84	140864	48.93	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.86%
64) 1,2-Dichlorobenzene-d4	11.86	152	107818	40.14	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.28%

Target Compounds					Ovalue
13) Acetone	2.34	43	55582	50.023 ug/L	100
15) Methyl Acetate	2.62	43	369923	192.152 ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091818\
Data File : VU026886.D
Acq On : 19 Sep 2018 00:32
Operator : MD/SY
Sample : J4865-20ME
Misc : 4.59µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DB3P8ME

Quant Time: Sep 19 08:17:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091718WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Sep 19 07:38:56 2018
Response via : Initial Calibration

