

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091819\
 Data File : VU034628.D
 Acq On : 17 Sep 2019 20:52
 Operator : JC/SP
 Sample : K4865-03
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COMJ9

Quant Time: Sep 18 08:23:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR091819WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Sep 18 08:05:17 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	98343	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	92621	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.46	152	38665	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	40750	5.21	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.20%
7) Chloroethane-d5	1.67	69	32503	5.12	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.40%
11) 1,1-Dichloroethene-d2	2.26	63	55209	4.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.00%
20) 2-Butanone-d5	4.17	46	98978	51.46	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.92%
24) Chloroform-d	4.62	84	63751	5.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.80%
26) 1,2-Dichloroethane-d4	5.29	65	34549	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.60%
32) Benzene-d6	5.31	84	130272	5.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.80%
36) 1,2-Dichloropropane-d6	6.31	67	44447	5.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.80%
41) Toluene-d8	7.54	98	119050	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.20%
43) trans-1,3-Dichloropropene-	7.83	79	16636	4.94	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.80%
46) 2-Hexanone-d5	8.30	63	78663	52.53	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.06%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	34601	5.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.20%
64) 1,2-Dichlorobenzene-d4	11.84	152	35790	5.37	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091819\
Data File : VU034628.D
Acq On : 17 Sep 2019 20:52
Operator : JC/SP
Sample : K4865-03
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
COMJ9

Quant Time: Sep 18 08:23:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR091819WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Sep 18 08:05:17 2019
Response via : Initial Calibration

