

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091820\
 Data File : VU040208.D
 Acq On : 18 Sep 2020 14:07
 Operator : SY/MD
 Sample : L4063-14
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AH9

Quant Time: Sep 19 00:15:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Sep 18 23:20:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	103144	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	102006	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	45218	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	29411	5.22	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.40%
7) Chloroethane-d5	1.92	69	27957	5.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.20%
11) 1,1-Dichloroethene-d2	2.58	63	51234	3.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.00%
20) 2-Butanone-d5	4.66	46	140871	62.54	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.08%
24) Chloroform-d	5.08	84	73636	5.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.00%
26) 1,2-Dichloroethane-d4	5.72	65	45934	5.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.20%
32) Benzene-d6	5.75	84	131330	5.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.20%
36) 1,2-Dichloropropane-d6	6.70	67	44484	5.67	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.40%
41) Toluene-d8	7.91	98	102746	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
43) trans-1,3-Dichloropropene-	8.19	79	18106	5.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.40%
46) 2-Hexanone-d5	8.64	63	100461	58.81	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	117.62%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	41630	6.13	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	122.60%#
64) 1,2-Dichlorobenzene-d4	12.19	152	42785	5.85	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.00%

Target Compounds

					Ovalue
13) Acetone	2.67	43	2832	1.358 ug/L	92
14) Carbon disulfide	2.81	76	2689	0.105 ug/L	99
34) Trichloroethene	6.55	95	956	0.098 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091820\
Data File : VU040208.D
Acq On : 18 Sep 2020 14:07
Operator : SY/MD
Sample : L4063-14
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AH9

Quant Time: Sep 19 00:15:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Sep 18 23:20:35 2020
Response via : Initial Calibration

