

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091823\
 Data File : VU055376.D
 Acq On : 18 Sep 2023 17:14
 Operator : MD/SY
 Sample : 04386-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AJ2

Quant Time: Sep 19 01:16:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091523WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Sep 19 01:12:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	248247	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	252987	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	80571	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	33402	2.341	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	46.800%	
7) Chloroethane-d5	1.911	69	46613	2.894	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	57.800%#	
11) 1,1-Dichloroethene-d2	2.563	65	11722	2.271	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	45.400%#	
20) 2-Butanone-d5	4.628	46	208528	52.850	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	105.700%	
24) Chloroform-d	5.058	84	118809	3.728	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	74.600%	
26) 1,2-Dichloroethane-d4	5.698	65	67099	4.173	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	83.400%	
32) Benzene-d6	5.724	84	185409	3.496	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	70.000%	
36) 1,2-Dichloropropane-d6	6.689	67	75086	4.155	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	83.200%	
41) Toluene-d8	7.898	98	126908	2.776	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	55.600%#	
43) trans-1,3-Dichloroprop...	8.181	79	24160	3.768	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	75.400%	
46) 2-Hexanone-d5	8.631	63	121720	52.178	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	104.360%	
56) 1,1,2,2-Tetrachloroeth...	10.753	84	54870	3.492	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	69.800%	
66) 1,2-Dichlorobenzene-d4	12.190	152	47164	4.546	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	91.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091823\
Data File : VU055376.D
Acq On : 18 Sep 2023 17:14
Operator : MD/SY
Sample : 04386-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AJ2

Quant Time: Sep 19 01:16:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091523WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Sep 19 01:12:51 2023
Response via : Initial Calibration

