

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091824\
 Data File : VU060871.D
 Acq On : 18 Sep 2024 15:37
 Operator : MD/SY
 Sample : P3973-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AS7

Quant Time: Sep 19 05:14:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 19 05:11:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	148267	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	159204	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	65188	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	41959	3.146	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.000%
7) Chloroethane-d5	1.911	69	37056	3.903	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.000%
11) 1,1-Dichloroethene-d2	2.560	65	16392	3.468	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.400%
20) 2-Butanone-d5	4.637	46	95320	44.524	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.040%
24) Chloroform-d	5.055	84	82786	4.061	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.200%
26) 1,2-Dichloroethane-d4	5.695	65	40745	4.181	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.600%
32) Benzene-d6	5.718	84	162093	3.799	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.000%
36) 1,2-Dichloropropane-d6	6.682	67	53757	4.072	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.400%
41) Toluene-d8	7.891	98	133608	3.434	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.600%#
43) trans-1,3-Dichloroprop...	8.177	79	18180	3.825	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.600%
46) 2-Hexanone-d5	8.631	63	73702	39.315	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.640%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	43926	4.247	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	49374	4.508	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091824\
Data File : VU060871.D
Acq On : 18 Sep 2024 15:37
Operator : MD/SY
Sample : P3973-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AS7

Quant Time: Sep 19 05:14:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 19 05:11:13 2024
Response via : Initial Calibration

