

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU091918\
 Data File : VU026892.D
 Acq On : 19 Sep 2018 11:38
 Operator : MD/SY
 Sample : VU0919MBL01
 Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK46

Quant Time: Sep 20 03:52:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091718WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Sep 20 02:30:06 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	220682	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	213601	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	86541	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	55928	34.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.52%
7) Chloroethane-d5	1.68	69	56446	40.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.80%
11) 1,1-Dichloroethene-d2	2.27	63	86895	28.13	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	56.26%#
21) 2-Butanone-d5	4.18	46	123331	113.84	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	113.84%
24) Chloroform-d	4.65	84	142680	47.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.24%
26) 1,2-Dichloroethane-d4	5.31	65	94772	49.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.40%
32) Benzene-d6	5.34	84	261530	43.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.38%
36) 1,2-Dichloropropane-d6	6.33	67	94980	49.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.48%
41) Toluene-d8	7.57	98	209895	38.81	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.62%#
43) trans-1,3-Dichloropropene-	7.85	79	37928	44.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.08%
47) 2-Hexanone-d5	8.31	63	82652	109.03	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.03%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	137603	51.17	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.34%
64) 1,2-Dichlorobenzene-d4	11.86	152	94386	52.43	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.86%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU091918\
Data File : VU026892.D
Acq On : 19 Sep 2018 11:38
Operator : MD/SY
Sample : VU0919MBL01
Misc : 5.00µ/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK46

Quant Time: Sep 20 03:52:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091718WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Sep 20 02:30:06 2018
Response via : Initial Calibration

