

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091923\
 Data File : VU055404.D
 Acq On : 19 Sep 2023 20:49
 Operator : MD/SY
 Sample : 04380-21
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH4P4

Quant Time: Sep 20 06:09:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 20 06:06:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	227411	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	233037	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	109994	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	33501	1.458	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	29.200%#
7) Chloroethane-d5	1.911	69	47333	2.337	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	46.800%#
11) 1,1-Dichloroethene-d2	2.567	65	19173	2.027	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	40.600%#
20) 2-Butanone-d5	4.631	46	177800	42.828	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.660%
24) Chloroform-d	5.058	84	128096	3.081	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	61.600%#
26) 1,2-Dichloroethane-d4	5.698	65	65925	3.297	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	66.000%#
32) Benzene-d6	5.724	84	231025	2.834	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	56.600%#
36) 1,2-Dichloropropane-d6	6.689	67	83387	3.424	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	68.400%
41) Toluene-d8	7.894	98	183052	2.464	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	49.200%#
43) trans-1,3-Dichloroprop...	8.181	79	14238	1.666	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	33.400%#
46) 2-Hexanone-d5	8.631	63	91258	35.104	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	70.200%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	67206	3.682	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	73.600%
66) 1,2-Dichlorobenzene-d4	12.193	152	70859	3.192	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	63.800%#
Target Compounds						Qvalue
3) Chloromethane	1.515	50	4194	0.215	ug/L	90
5) Vinyl chloride	1.602	62	138648	6.598	ug/L	99
12) 1,1-Dichloroethene	2.576	96	76632	4.820	ug/L #	81
16) Methylene chloride	3.039	84	7073	0.371	ug/L	92
17) Methyl tert-butyl Ether	3.361	73	16843	0.488	ug/L	96
18) trans-1,2-Dichloroethene	3.354	96	6071	0.379	ug/L	95
19) 1,1-Dichloroethane	3.866	63	47384	1.599	ug/L	98
22) cis-1,2-Dichloroethene	4.660	96	1385956	74.593	ug/L	100
27) 1,2-Dichloroethane	5.798	62	11582	0.602	ug/L #	94
34) Trichloroethene	6.537	95	1112601	58.643	ug/L	98
47) Tetrachloroethene	8.550	164	9941	0.730	ug/L	97
51) Chlorobenzene	9.447	112	13063	0.285	ug/L #	83
64) 1,3-Dichlorobenzene	11.746	146	3597	0.119	ug/L #	91
65) 1,4-Dichlorobenzene	11.833	146	5834	0.192	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091923\
Data File : VU055404.D
Acq On : 19 Sep 2023 20:49
Operator : MD/SY
Sample : 04380-21
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH4P4

Quant Time: Sep 20 06:09:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 20 06:06:30 2023
Response via : Initial Calibration

