

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091923\
 Data File : VU055400.D
 Acq On : 19 Sep 2023 19:12
 Operator : MD/SY
 Sample : 04386-20
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Sep 20 06:08:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Sep 20 06:06:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	234830	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	247264	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.810	152	120942	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	115971	4.886	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.800%
7) Chloroethane-d5	1.911	69	103539	4.950	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.000%
11) 1,1-Dichloroethene-d2	2.563	65	48924	5.008	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.200%
20) 2-Butanone-d5	4.631	46	228755	53.361	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.720%
24) Chloroform-d	5.058	84	218391	5.087	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.800%
26) 1,2-Dichloroethane-d4	5.698	65	110308	5.342	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.800%
32) Benzene-d6	5.724	84	445600	5.152	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.000%
36) 1,2-Dichloropropane-d6	6.688	67	135085	5.227	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.600%
41) Toluene-d8	7.894	98	386215	4.899	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.000%
43) trans-1,3-Dichloroprop...	8.177	79	45381	5.004	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.000%
46) 2-Hexanone-d5	8.631	63	129821	47.064	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.120%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	97681	5.044	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	132878	5.444	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091923\
Data File : VU055400.D
Acq On : 19 Sep 2023 19:12
Operator : MD/SY
Sample : 04386-20
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Sep 20 06:08:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Sep 20 06:06:30 2023
Response via : Initial Calibration

