

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091924\
 Data File : VU060892.D
 Acq On : 19 Sep 2024 12:45
 Operator : MD/SY
 Sample : P3952-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B74

Quant Time: Sep 19 23:56:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Sep 19 23:54:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	128275	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	135934	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	54079	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	39228	3.400	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	68.000%	
7) Chloroethane-d5	1.911	69	33775	4.112	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	82.200%	
11) 1,1-Dichloroethene-d2	2.563	65	16063	3.928	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	78.600%	
20) 2-Butanone-d5	4.637	46	93409	50.432	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	100.860%	
24) Chloroform-d	5.055	84	80732	4.577	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	91.600%	
26) 1,2-Dichloroethane-d4	5.695	65	42399	5.029	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	100.600%	
32) Benzene-d6	5.721	84	150942	4.143	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	82.800%	
36) 1,2-Dichloropropane-d6	6.682	67	50776	4.504	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	90.000%	
41) Toluene-d8	7.891	98	126370	3.804	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	76.000%	
43) trans-1,3-Dichloroprop...	8.177	79	16796	4.139	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	82.800%	
46) 2-Hexanone-d5	8.630	63	72740	45.444	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.880%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	40303	4.564	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	91.200%	
66) 1,2-Dichlorobenzene-d4	12.187	152	47159	5.190	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091924\
Data File : VU060892.D
Acq On : 19 Sep 2024 12:45
Operator : MD/SY
Sample : P3952-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B74

Quant Time: Sep 19 23:56:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Sep 19 23:54:48 2024
Response via : Initial Calibration

