

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091924\
 Data File : VU060925.D
 Acq On : 20 Sep 2024 04:39
 Operator : MD/SY
 Sample : P4092-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B78

Quant Time: Sep 20 05:58:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 20 01:34:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.241	114	123248	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	133186	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	54647	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	29434	2.655	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	53.200%
7) Chloroethane-d5	1.910	69	27898	3.535	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.600%
11) 1,1-Dichloroethene-d2	2.560	65	12810	3.260	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.200%
20) 2-Butanone-d5	4.634	46	78430	44.072	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.140%
24) Chloroform-d	5.055	84	73225	4.321	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.400%
26) 1,2-Dichloroethane-d4	5.695	65	36827	4.546	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.000%
32) Benzene-d6	5.721	84	126954	3.556	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	71.200%
36) 1,2-Dichloropropane-d6	6.682	67	42371	3.836	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	76.800%
41) Toluene-d8	7.891	98	102664	3.154	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	63.000%#
43) trans-1,3-Dichloroprop...	8.174	79	13682	3.441	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	68.800%
46) 2-Hexanone-d5	8.627	63	66055	42.119	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.240%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	39061	4.514	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	42531	4.632	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU091924\
Data File : VU060925.D
Acq On : 20 Sep 2024 04:39
Operator : MD/SY
Sample : P4092-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B78

Quant Time: Sep 20 05:58:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091624WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 20 01:34:14 2024
Response via : Initial Calibration

