

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092018\
 Data File : VU026905.D
 Acq On : 19 Sep 2018 18:29
 Operator : MD/SY
 Sample : J4913-16
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BPS1-TT-MW302D-20180913

Quant Time: Sep 20 06:54:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U091218W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 05:32:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	81527	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	139068	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	136919	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	73089	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	68671	55.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.06%	
35) Dibromofluoromethane	4.89	113	62523	55.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.24%	
50) Toluene-d8	7.57	98	204688	48.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.58%	
62) 4-Bromofluorobenzene	10.31	95	68663	45.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.32%	
Target Compounds						
16) Acetone	2.32	43	1699	2.426	ug/l	91
44) Trichloroethene	6.19	130	2037	1.697	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092018\
Data File : VU026905.D
Acq On : 19 Sep 2018 18:29
Operator : MD/SY
Sample : J4913-16
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BPS1-TT-MW302D-20180913

Quant Time: Sep 20 06:54:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U091218W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 05:32:45 2018
Response via : Initial Calibration

