

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092018\
 Data File : VU026911.D
 Acq On : 19 Sep 2018 20:48
 Operator : MD/SY
 Sample : J4934-05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-MW132I-20180912

Quant Time: Sep 20 07:18:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U091218W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 12 05:32:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	80714	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	136139	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	136110	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	72671	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	69593	56.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.68%	
35) Dibromofluoromethane	4.88	113	61849	55.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.40%	
50) Toluene-d8	7.57	98	200347	48.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.56%	
62) 4-Bromofluorobenzene	10.31	95	68106	45.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.50%	
Target Compounds						
12) 1,1-Dichloroethene	2.28	96	2074	2.238	ug/l	98
16) Acetone	2.32	43	2015	2.907	ug/l	99
44) Trichloroethene	6.20	130	1401	1.193	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092018\
Data File : VU026911.D
Acq On : 19 Sep 2018 20:48
Operator : MD/SY
Sample : J4934-05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-MW132I-20180912

Quant Time: Sep 20 07:18:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U091218W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 12 05:32:45 2018
Response via : Initial Calibration

