

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092019\
 Data File : VU034719.D
 Acq On : 21 Sep 2019 03:27
 Operator : JC/SP
 Sample : K5001-07
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFCY2

Quant Time: Sep 21 06:51:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091919WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Sep 21 05:27:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	461946	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	440819	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.46	152	191387	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	162448	34.35	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.70%
7) Chloroethane-d5	1.67	69	142427	38.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	77.94%
11) 1,1-Dichloroethene-d2	2.26	63	231064	28.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.70%#
21) 2-Butanone-d5	4.15	46	346012	95.47	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.47%
24) Chloroform-d	4.62	84	300555	43.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.44%
26) 1,2-Dichloroethane-d4	5.29	65	221734	43.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.74%
32) Benzene-d6	5.32	84	601182	44.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.00%
36) 1,2-Dichloropropane-d6	6.31	67	212979	44.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.20%
41) Toluene-d8	7.54	98	556618	42.82	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.64%
43) trans-1,3-Dichloropropene-	7.83	79	89230	39.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.02%
47) 2-Hexanone-d5	8.29	63	203873	85.00	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.00%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	288051	43.47	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.94%
64) 1,2-Dichlorobenzene-d4	11.84	152	180507	44.98	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.96%

Target Compounds					Ovalue
13) Acetone	2.31	43	21486	5.096 ug/L	99
46) Tetrachloroethene	8.21	164	8252	3.107 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092019\
Data File : VU034719.D
Acq On : 21 Sep 2019 03:27
Operator : JC/SP
Sample : K5001-07
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFCY2

Quant Time: Sep 21 06:51:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091919WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Sep 21 05:27:34 2019
Response via : Initial Calibration

