

Data File : VU026938.D
Acq On : 20 Sep 2018 15:02
Operator : MD/SY
Sample : J4893-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C08L1

Manual Integrations
APPROVED

sam
9/28/2018 4:05:36 PM

Quant Time: Sep 28 05:48:19 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091718WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 21 04:24:07 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	285597	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.10	117	284789	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	173590	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	76104	36.02	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.04%
7) Chloroethane-d5	1.68	69	64756	35.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	71.62%
11) 1,1-Dichloroethene-d2	2.27	63	107240	26.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	53.66%#
21) 2-Butanone-d5	4.18	46	103830	74.06	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	74.06%
24) Chloroform-d	4.65	84	129075	32.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	65.88%#
26) 1,2-Dichloroethane-d4	5.31	65	83550	33.86	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	67.72%#
32) Benzene-d6	5.34	84	270282	33.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	67.74%#
36) 1,2-Dichloropropane-d6	6.33	67	85676	33.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	66.64%#
41) Toluene-d8	7.57	98	243279	33.74	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	67.48%#
43) trans-1,3-Dichloropropene-	7.85	79	38809	33.80	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.60%
47) 2-Hexanone-d5	8.31	63	74126	73.34	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	73.34%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	116511	32.49	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	64.98%#
64) 1,2-Dichlorobenzene-d4	11.88	152	128463	35.57	ug/L	0.02
Spiked Amount	50.000	Range	80 - 120	Recovery	=	71.14%#

Target Compounds

					Qvalue
13) Acetone	2.32	43	3959	2.873 ug/L	90
29) Cyclohexane	5.00	56	5210	1.576 ug/L	98
31) Carbon tetrachloride	5.13	117	3998	1.317 ug/L	94
33) Benzene	5.39	78	128235	14.523 ug/L	100
35) Methylcyclohexane	6.42	83	18807	5.445 ug/L #	78
46) Tetrachloroethene	8.23	164	1952	1.039 ug/L	88
51) Chlorobenzene	9.13	112	17312209m	2855.443 ug/L	
62) 1,3-Dichlorobenzene	11.42	146	2452161	463.387 ug/L	99
63) 1,4-Dichlorobenzene	11.52	146	14927476	2561.775 ug/L	96
65) 1,2-Dichlorobenzene	11.89	146	24660266m	4208.626 ug/L	
67) 1,3,5-Trichlorobenzene	12.89	180	10525	2.570 ug/L	99
68) 1,2,4-trichlorobenzene	13.52	180	16277738m	4612.801 ug/L	
70) 1,2,3-Trichlorobenzene	13.99	180	2353011	605.509 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VU026938.D
Acq On : 20 Sep 2018 15:02
Operator : MD/SY
Sample : J4893-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C08L1

Manual Integrations
APPROVED

sam
9/28/2018 4:05:36 PM

Quant Time: Sep 28 05:48:19 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM091718WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Sep 21 04:24:07 2018
Response via : Initial Calibration

