

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092120\
 Data File : VU040241.D
 Acq On : 21 Sep 2020 18:08
 Operator : SY/MD
 Sample : VIBLK62
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK62

Quant Time: Sep 22 08:58:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 22 07:07:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	116550	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	115732	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	51667	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	36140	5.68	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	113.60%
7) Chloroethane-d5	1.93	69	32983	6.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	124.40%
11) 1,1-Dichloroethene-d2	2.58	63	63138	3.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.40%
20) 2-Butanone-d5	4.66	46	161517	63.46	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	126.92%
24) Chloroform-d	5.08	84	79384	5.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.80%
26) 1,2-Dichloroethane-d4	5.72	65	51133	5.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.40%
32) Benzene-d6	5.75	84	151889	5.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.40%
36) 1,2-Dichloropropane-d6	6.70	67	50229	5.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	112.80%
41) Toluene-d8	7.91	98	128729	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.40%
43) trans-1,3-Dichloropropene-	8.19	79	19851	4.85	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.00%
46) 2-Hexanone-d5	8.64	63	119304	61.56	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	123.12%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	43937	5.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	114.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	50873	6.08	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	121.60%#

Target Compounds					Ovalue
16) Methylene chloride	3.07	84	2578	0.218 ug/L	96
51) Chlorobenzene	9.45	112	3829	0.145 ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092120\
Data File : VU040241.D
Acq On : 21 Sep 2020 18:08
Operator : SY/MD
Sample : VIBLK62
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK62

Quant Time: Sep 22 08:58:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 22 07:07:10 2020
Response via : Initial Calibration

