

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092120\
 Data File : VU040244.D
 Acq On : 21 Sep 2020 19:18
 Operator : SY/MD
 Sample : L4063-11
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Sep 22 09:06:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Sep 22 07:07:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	117512	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	115951	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	51791	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	35501	5.53	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.60%
7) Chloroethane-d5	1.92	69	32729	6.12	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	122.40%
11) 1,1-Dichloroethene-d2	2.58	63	61242	3.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.40%
20) 2-Butanone-d5	4.66	46	158462	61.75	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.50%
24) Chloroform-d	5.08	84	78820	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.20%
26) 1,2-Dichloroethane-d4	5.72	65	52063	5.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.60%
32) Benzene-d6	5.74	84	151986	5.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.20%
36) 1,2-Dichloropropane-d6	6.70	67	49131	5.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.00%
41) Toluene-d8	7.90	98	128308	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.00%
43) trans-1,3-Dichloropropene-	8.19	79	20107	4.91	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.20%
46) 2-Hexanone-d5	8.64	63	119506	61.55	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	123.10%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	45506	5.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	118.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	50046	5.97	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.40%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	1761	0.148 ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092120\
Data File : VU040244.D
Acq On : 21 Sep 2020 19:18
Operator : SY/MD
Sample : L4063-11
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Sep 22 09:06:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR083120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Sep 22 07:07:10 2020
Response via : Initial Calibration

