

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092123\
 Data File : VU055441.D
 Acq On : 21 Sep 2023 12:08
 Operator : MD/SY
 Sample : 04398-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Sep 22 02:33:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Sep 22 02:32:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.248	114	257722	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	265309	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	126191	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	112601	4.323	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.400%
7) Chloroethane-d5	1.911	69	104411	4.549	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.000%
11) 1,1-Dichloroethene-d2	2.563	65	49394	4.607	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.200%
20) 2-Butanone-d5	4.634	46	239777	50.964	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.920%
24) Chloroform-d	5.058	84	236989	5.030	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.600%
26) 1,2-Dichloroethane-d4	5.698	65	122406	5.401	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.000%
32) Benzene-d6	5.724	84	465618	5.017	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.400%
36) 1,2-Dichloropropane-d6	6.689	67	143319	5.169	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.400%
41) Toluene-d8	7.894	98	401726	4.750	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.000%
43) trans-1,3-Dichloroprop...	8.177	79	47760	4.908	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.200%
46) 2-Hexanone-d5	8.634	63	125569	42.427	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.860%
56) 1,1,2,2-Tetrachloroeth...	10.756	84	100953	4.858	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	135663	5.326	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU092123\
Data File : VU055441.D
Acq On : 21 Sep 2023 12:08
Operator : MD/SY
Sample : 04398-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Sep 22 02:33:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR091923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Sep 22 02:32:08 2023
Response via : Initial Calibration

