

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092218\
 Data File : VU026989.D
 Acq On : 21 Sep 2018 23:58
 Operator : MD/SY
 Sample : J4916-19
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-PZ13811-20180911

Quant Time: Sep 22 04:13:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	72473	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	145931	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	151602	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	73074	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	86973	59.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	119.26%	
35) Dibromofluoromethane	4.89	113	65598	51.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.72%	
50) Toluene-d8	7.57	98	238135	49.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.80%	
62) 4-Bromofluorobenzene	10.31	95	81051	49.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.00%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092218\
Data File : VU026989.D
Acq On : 21 Sep 2018 23:58
Operator : MD/SY
Sample : J4916-19
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-PZ138I1-20180911

Quant Time: Sep 22 04:13:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

