

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092218\
 Data File : VU026994.D
 Acq On : 22 Sep 2018 04:12
 Operator : MD/SY
 Sample : VU0922WBL02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0922WBL02

Quant Time: Sep 22 04:54:45 2018
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	76653	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	151694	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	153217	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	73918	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	91048	59.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.04%	
35) Dibromofluoromethane	4.88	113	67935	51.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.34%	
50) Toluene-d8	7.57	98	237855	47.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.94%	
62) 4-Bromofluorobenzene	10.31	95	83629	48.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.28%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092218\
Data File : VU026994.D
Acq On : 22 Sep 2018 04:12
Operator : MD/SY
Sample : VU0922WBL02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0922WBL02

Quant Time: Sep 22 04:54:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

