

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092218\
 Data File : VU027004.D
 Acq On : 22 Sep 2018 08:04
 Operator : MD/SY
 Sample : J4934-22
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-PZ15711-20180913

Quant Time: Sep 24 04:06:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	73832	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	146767	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	150092	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	72690	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	89425	60.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	120.36%	
35) Dibromofluoromethane	4.89	113	66986	52.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.30%	
50) Toluene-d8	7.57	98	242270	49.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.94%	
62) 4-Bromofluorobenzene	10.31	95	81642	49.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.16%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092218\
Data File : VU027004.D
Acq On : 22 Sep 2018 08:04
Operator : MD/SY
Sample : J4934-22
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-PZ15711-20180913

Quant Time: Sep 24 04:06:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

