

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU092218\
 Data File : VU027033.D
 Acq On : 22 Sep 2018 22:00
 Operator : MD/SY
 Sample : J4916-16
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 73 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 SA-PZ139I-20180911

Quant Time: Sep 24 05:32:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	91884	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	153917	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	150523	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	70139	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	95117	51.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.86%	
35) Dibromofluoromethane	4.88	113	65616	49.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.36%	
50) Toluene-d8	7.57	98	239740	47.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.30%	
62) 4-Bromofluorobenzene	10.31	95	79634	45.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.28%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU092218\
Data File : VU027033.D
Acq On : 22 Sep 2018 22:00
Operator : MD/SY
Sample : J4916-16
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 73 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
SA-PZ139I-20180911

Quant Time: Sep 24 05:32:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

