

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU092218\
 Data File : VU027039.D
 Acq On : 23 Sep 2018 00:19
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 79 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 24 05:41:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.653	0.637	2.5	95	0.00
3 P	Chloromethane	1.126	1.061	5.8	96	0.00
4 C	Vinyl Chloride	1.051	1.032	1.8#	96	0.00
5 T	Bromomethane	0.598	0.554	7.4	98	0.00
6 T	Chloroethane	0.671	0.667	0.6	98	0.00
7 T	Trichlorofluoromethane	1.193	1.174	1.6	96	0.00
8 T	Diethyl Ether	0.574	0.576	-0.3	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.734	0.706	3.8	96	0.00
10 T	Methyl Iodide	0.910	0.803	11.8	84	0.00
11 T	Tert butyl alcohol	0.231	0.245	-6.1	104	0.00
12 CM	1,1-Dichloroethene	0.691	0.677	2.0#	97	0.00
13 T	Acrolein	0.124	0.123	0.8	90	0.00
14 T	Allyl chloride	1.401	1.523	-8.7	104	0.00
15 T	Acrylonitrile	0.581	0.611	-5.2	99	0.00
16 T	Acetone	0.658	0.539	18.1	93	0.00
17 T	Carbon Disulfide	2.250	2.196	2.4	97	0.00
18 T	Methyl Acetate	1.432	1.530	-6.8	102	0.00
19 T	Methyl tert-butyl Ether	2.538	2.766	-9.0	103	0.00
20 T	Methylene Chloride	0.901	0.863	4.2	98	0.00
21 T	trans-1,2-Dichloroethene	0.737	0.761	-3.3	97	0.00
22 T	Diisopropyl ether	2.996	3.394	-13.3	103	0.00
23 T	Vinyl Acetate	2.554	2.924	-14.5	104	0.00
24 P	1,1-Dichloroethane	1.650	1.753	-6.2	102	0.00
25 T	2-Butanone	0.846	0.869	-2.7	99	0.00
26 T	2,2-Dichloropropane	1.287	1.129	12.3	85	0.00
27 T	cis-1,2-Dichloroethene	0.833	0.871	-4.6	99	0.00
28 T	Bromochloromethane	0.795	0.894	-12.5	104	0.00
29 T	Tetrahydrofuran	0.512	0.568	-10.9	102	0.00
30 C	Chloroform	1.532	1.550	-1.2#	98	0.00
31 T	Cyclohexane	1.687	1.569	7.0	100	0.00
32 T	1,1,1-Trichloroethane	1.223	1.278	-4.5	99	0.00
33 S	1,2-Dichloroethane-d4	1.006	1.059	-5.3	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.433	0.426	1.6	97	0.00
36 T	1,1-Dichloropropene	0.667	0.683	-2.4	102	0.00
37 T	Ethyl Acetate	0.903	0.927	-2.7	102	0.00
38 T	Carbon Tetrachloride	0.635	0.624	1.7	97	0.00
39 T	Methylcyclohexane	0.725	0.769	-6.1	98	0.00
40 TM	Benzene	2.029	2.016	0.6	99	0.00
41 T	Methacrylonitrile	0.469	0.524	-11.7	104	0.00
42 TM	1,2-Dichloroethane	0.766	0.780	-1.8	100	0.00
43 T	Isopropyl Acetate	1.317	1.430	-8.6	104	0.00
44 TM	Trichloroethene	0.432	0.430	0.5	98	0.00
45 C	1,2-Dichloropropane	0.578	0.593	-2.6#	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU092218\
 Data File : VU027039.D
 Acq On : 23 Sep 2018 00:19
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 79 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 24 05:41:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 22 01:31:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.349	0.345	1.1	98	0.00
47 T	Bromodichloromethane	0.706	0.703	0.4	100	0.00
48 T	Methyl methacrylate	0.642	0.723	-12.6	103	0.00
49 T	1,4-Dioxane	0.013	0.014	-7.7	98	0.00
50 S	Toluene-d8	1.652	1.661	-0.5	97	0.00
51 T	4-Methyl-2-Pentanone	0.886	0.971	-9.6	103	0.00
52 CM	Toluene	1.163	1.225	-5.3#	97	0.00
53 T	t-1,3-Dichloropropene	0.758	0.815	-7.5	101	0.00
54 T	cis-1,3-Dichloropropene	0.834	0.879	-5.4	99	0.00
55 T	1,1,2-Trichloroethane	0.493	0.500	-1.4	98	0.00
56 T	Ethyl methacrylate	0.733	0.849	-15.8	105	0.00
57 T	1,3-Dichloropropane	0.916	0.958	-4.6	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.420	0.479	-14.0	99	0.00
59 T	2-Hexanone	0.694	0.745	-7.3	102	0.00
60 T	Dibromochloromethane	0.483	0.500	-3.5	100	0.00
61 T	1,2-Dibromoethane	0.513	0.521	-1.6	98	0.00
62 S	4-Bromofluorobenzene	0.567	0.603	-6.3	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.385	0.382	0.8	99	0.00
65 PM	Chlorobenzene	1.326	1.302	1.8	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.457	0.465	-1.8	99	0.00
67 C	Ethyl Benzene	2.180	2.367	-8.6#	100	0.00
68 T	m/p-Xylenes	0.798	0.884	-10.8	98	0.00
69 T	o-Xylene	0.772	0.843	-9.2	99	0.00
70 T	Styrene	1.247	1.415	-13.5	99	0.00
71 P	Bromoform	0.353	0.370	-4.8	101	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.803	4.220	-11.0	100	0.00
74 T	N-amyl acetate	2.202	2.454	-11.4	104	0.00
75 P	1,1,2,2-Tetrachloroethane	1.749	1.647	5.8	97	0.00
76 T	1,2,3-Trichloropropane	1.519	1.411	7.1	92	0.00
77 T	Bromobenzene	0.929	0.946	-1.8	97	0.00
78 T	n-propylbenzene	4.500	5.080	-12.9	98	0.00
79 T	2-Chlorotoluene	2.827	3.009	-6.4	99	0.00
80 T	1,3,5-Trimethylbenzene	3.140	3.520	-12.1	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.475	0.456	4.0	96	0.00
82 T	4-Chlorotoluene	3.170	3.477	-9.7	99	0.00
83 T	tert-Butylbenzene	3.104	3.354	-8.1	99	0.00
84 T	1,2,4-Trimethylbenzene	3.195	3.626	-13.5	99	0.00
85 T	sec-Butylbenzene	3.789	4.266	-12.6	99	0.00
86 T	p-Isopropyltoluene	3.186	3.631	-14.0	98	0.00
87 T	1,3-Dichlorobenzene	1.730	1.746	-0.9	97	0.00
88 T	1,4-Dichlorobenzene	1.885	1.790	5.0	97	0.00
89 T	n-Butylbenzene	2.969	3.383	-13.9	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU092218\
Data File : VU027039.D
Acq On : 23 Sep 2018 00:19
Operator : MD/SY
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 79 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Sep 24 05:41:25 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U092218W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 22 01:31:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.689	0.678	1.6	99	0.00
91 T	1,2-Dichlorobenzene	1.837	1.775	3.4	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.357	0.368	-3.1	101	0.00
93 T	1,2,4-Trichlorobenzene	1.059	1.155	-9.1	99	0.00
94 T	Hexachlorobutadiene	0.560	0.555	0.9	95	0.00
95 T	Naphthalene	3.647	4.193	-15.0	98	0.00
96 T	1,2,3-Trichlorobenzene	1.135	1.210	-6.6	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6